DECLARATION OF CONFORMITY

MCD Medical Computers Deutschland GmbH

Konrad-Zuse-Ring 17 A/B

41179 Moenchengladbach

- **Germany –**

declares under sole responsibility that the system

Modell: 5200159M	**MCD Medical Line OMNI.view 21.5”**
5200160M	**MCD Medical Line OMNI.view 21.5” T**

with the essential requirements of the following EC Directives, if used for its intended purpose:

EMC Directive 2004/108/EG

Low Voltage Directive 2006/95/EG

Applied harmonized standards or normative documents:

EN61000-4-2:2009 / IEC 61000-4-2:2008
EN61000-4-3:2006 + A1:2008 + A2:2010 / IEC 61000-4-3:2006 A1:2007 + A2:2012
EN61000-4-8:2010 / IEC 61000-4-8:2009
IEC 60601-1-2:2014

Power supply Magic Power Technology MPM-X60:
EN 60601-1 3rd Edition, ANSI/AAMI ES60601-1 3rd Edition, CAN/CSA-C22.2 No. 60601-1:08

This declaration certifies compliance with the indicated directives but implies no warranty of characteristics. The safety of the supplied product documentation must be observed.

Moenchengladbach, 2016-01-29	Thomas Hollex	
(Place and date of the exhibit)	(Issuer)	(Signature)

Management System EN ISO 13485:2012

www.tuv.com ID 9105055459